

EU Type Examination Certificate

No. 0200-NAWI-07541

BX65 / BX66

NON-AUTOMATIC WEIGHING INDICATOR

Issued by **FORCE Certification**
EU - Notified Body No. 0200

In accordance with the requirements for the non-automatic weighing instrument of the EU Council Directive 2014/31/EU.

Issued to **BAYKON Endüstriyel Kontrol Sistemleri San ve Tic A.S.**
Tuzla Kimya Sanayicileri OSB
Organik Caddesi 31
Tepeören, Tuzla
34956 Istanbul
TURKEY

In respect of Non-automatic weighing instrument designated BX65 / BX66 with variants of modules of load receptors, load cells and peripheral equipment.
Accuracy class : III and IIII
Maximum capacity, Max: From 1 kg up to 300 000 kg
Verification scale interval: $e_i = \text{Max}_i / n_i$
Maximum number of verification scale intervals: $n_i \leq 10000$ for single-interval / multi-range / multi-interval (however, dependent on environment and the composition of the modules).
Variants of modules and conditions for the composition of the modules are set out in the annex.

The conformity with the essential requirements in annex 1 of the Directive is met by the application of the European Standard EN 45501:2015 and OIML R76:2006.

Note: This certificate is a revised edition of DK0199.621.

The principal characteristics and approval conditions are set out in the descriptive annex to this certificate.

The annex comprises 18 pages.

Issued on **2019-11-07**
Valid until **2026-12-15**

FORCE Certification references:

Task no.: 119-35172.90.10 and ID no.: 0200-NAWI-07541

Signatory: J. Hovgård Jensen