



tel. +420 545 555 111, fax +420 545 222 728, www.cmi.cz

## EC TYPE-APPROVAL CERTIFICATE

Number: TCM 128/14 - 5180

Page 1 from 3 pages

In accordance:

with Directive 2009/23/EC of the European Parliament and of the Council as amended – codified version of Directive 90/384/EEC implemented by Government Order No. 326/2002 Coll. as amended, that lays down technical requirements on non-automatic weighing

instruments.

Manufacturer:

**BAYKON Endustriyel Kontrol Sistemleri A.S.** 

Tuzla Kimyacilar Organize Sanayi Bolgesi Organik Cad. No:31, Tuzla, Istanbul - Turkey

For:

non-automatic weighing instrument

Type: BX 61 ...

Accuracy class (III) or (III)

Max according the module composition

 $n \le 6000$  for class (III) or  $n \le 1000$  for class (III)

Temperature range -10°C/+40°C

Valid until:

26 March 2024 7

**Document No:** 

0511-CS-A023-14

Description:

Essential characteristics, approved conditions and special conditions, if any, are described in

this certificate.

Date of issue:

27 March 2014

Certificate approved by:

RNDr. Pavel Klenovský

This certificate was issued according to EC type-examination according to point 1 of annex II to Directive 2009/23/EC of the European Parliament and of the Council - codified version of 90/384/EEC or point 1 of annex 2 to Government Order 326/2002 Coll., respectively.